



Nimagen B.V.
P.O. Box 91 | 6500 AB Nijmegen | The Netherlands
Tel. +31 (0)24 820 0241 | Fax. +31 (0)24 358 0259
info@nimagen.com
VAT#: NL850011243B01
Rabobank Nijmegen: NL62RABO 01 55 82 47 83

CERTIFICATE OF ANALYSIS

PRODUCT	BrilliantDye™ Terminator (v1.1) Cycle Sequencing kit
PART NO.	BRD1-XXX
LOT NO.	604436150
EXPIRATION DATE	15-DEC-2022
APPEARANCE	Containers are inspected for correct label, quantity and lot number. Caps are tight with no signs of leakage. Clear pink liquid, aqueous solution.
ANALYSIS¹	RESULT
QV20+ Report	PASS
Contiguous Read Length	PASS
Trace Score	PASS
Base calling accuracy $\geq 98.0\%$	PASS
Fluorescence background	PASS

Quality Assurance Manager: Dirk Franssen

Date: 14-6-2021

¹ 1 μ L PGEM control from kit with 1 μ L -21M13 primer from kit is sequenced according to standard protocol, with 1 μ L BrilliantDye Terminator v1.1 Sequencing mix, 1.5 μ L 5x Sequencing Buffer and 5.5 μ L water (total volume = 10 μ L), purified and analyzed on a 3130 Genetic Analyzer.